

**Substances that elicit positive and false-positive results using
Siemens Emit® reagents**

Emit® II Plus Buprenorphine Assay (5 ng cutoff) Calibrated with Buprenorphine at 5 ng/ml

Drug	ng/mL
Norbuprenorphine	5
Buprenorphine Glucuronide	1000
Norbuprenorphine Glucuronide	1000

False Positives: None Listed